



广 微 测
Gmicro Testing

GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS

Report №.

2020SP8367R01E

Name of Sample

Air sterilizer AP-120

Applicant

ETS Products Philippines Inc

Test Type

Entrustment Test



Address: Building 66, No.100 Central Xian Lie Road, Guangzhou, China

Postcode: 510070

Tel: +86 20 87137666

Fax : +86 20 87137668

Website : www.gddcm.com




GUANGDONG DETECTION CENTER OF MICROBIOLOGY

REPORT FOR ANALYSIS



Report №.: 2020SP8367R01E Verification Code: 45082391

Name of Sample	Air sterilizer AP-120	Test Type	Entrustment Test
Applicant	ETS Products Philippines Inc	Address	Berthaphil iV, 2C 2000 Clark Freeport Zone Civil Aviation Complex Philippines
Sample Source	Submitted for Testing by the Applicant	Sample Quantity	One
Spec and Lot № of Sample	AP-120 Lot number: 90092007230015 Specification: 220V 50HZ	State and Characteristic	White machine
Sample Received Date	2020-08-05	Test Completion Date	2020-08-27
Test Standard and Method	Refer to GB 21551.3-2010 Appendix A		
Item Tested	Air virus removal effect identification test		
Test Conclusion	The test data of the sample(s) is attached to the page(s) of this report.		
Remarks	<p>Issue Date: 2020-09-04 (Official Seal)</p> 		

生物
检验

Editor: Chen Jingxun

Verifier: Liang Daqing

Approver: Li Liangqiu



GUANGDONG DETECTION CENTER OF MICROBIOLOGY

ANALYSIS AND TEST RESULT

Report No.: 2020SP8367R01E

- 1. Test space: 30 m³ Test cabinet
- 2. Action time: 2h (Sample volume is 2mL)
- 3. Test results:

Test virus & host cell	Test number	Air virus content (TCID ₅₀ /m ³)		Removal rate (%)
		Before test	After the test	
Influenza A virus H1N1 host cell: MDCK cells	1	3.08×10 ⁶	1.44×10 ²	99.98
	2	1.94×10 ⁶	1.14×10 ²	99.98
	3	4.88×10 ⁶	1.44×10 ²	99.99
Influenza A virus H3N2 host cell: MDCK cells	1	4.06×10 ⁶	4.88×10 ²	99.95
	2	3.08×10 ⁶	2.44×10 ²	99.97
	3	3.08×10 ⁶	1.14×10 ²	99.98

Note :The result of bactericidal rate test has eliminated the effect of microbial natural decay in air.
Cell control: Normal growth.

Photo of the sample:



(Blank below)

分析
专用



Report №.: 2020SP8367R01E

Notice Items

1. The Test report is invalid if not affixed with Authorized Stamp of Test and Paging Seal.
2. The Test report is invalid without signature of verifier and approver.
3. The Test report is invalid if being supplemented, deleted or altered.
4. Without prior written permission, the report cannot be reproduced, except in full.
5. Unless otherwise stated, the results shown in this test report refer only to the sample(s) submitted.
6. Any dispute of the report must be raised to the testing body within 15 days after the report is received, exceeding which the dispute will not be accepted.
7. For the tested sample(s) submitted by the applicant, the sample information in the test report is provided by the applicant and the laboratory is not responsible for its authenticity.
8. This test report is for reference only to the applicant and does not have a proof of effect for others.

